

# Anabolic Lab™

Analytical Lab Testing for  
Anabolic Steroid Harm Reduction

<https://anaboliclab.com>

Lab Report – April 14, 2016

QD Labs  
Trenbolone Enanthate (200mg/mL)  
Trenbolone Enanthate



If you like the attached report, please visit [our website](https://anaboliclab.com) to find out how you can support independent and objective lab testing.



Simec AG

Areal Bleiche West Postfach 413 CH - 4800 Zofingen  
Telefon +41 62 752 83 08 Fax +41 62 752 83 09  
E-Mail: info@simec.ch www.simec.ch

# Test report

Number **C 42745**

**Object of analysis:** QDLabs Trenbolone enanthate 200 mg/mL

**Customer:** AnabolicLab.com **Samples:** 1

**Batch:** 1362522, Exp. 09/2017 **Receipt:** 01.04.2016

**Subject:** Assay, Identification Trenbolone enanthate

**Method(s):** HPLC

Sample / Analysis:	Method	Status	Result	Specification	complies
<b><u>Trenbolone enanthate:</u></b>					
Identification	HPLC	N	complies	complies	yes
Content	HPLC	N	241.3 mg/mL	200 mg/mL	-

**Method-status:** G = GMP A = accredited V = generally validated P = validated on product N = not validated  
E = external lab

**Remarks:** The sample(s) mentioned above have been analysed as they have been sent to us by the client. We have not controlled external processes like manufacturing, labelling, sampling, shipping, storage. These results are for information only and do not compensate for correct quality control by the manufacturer/distributor. Generally, pharmaceutical products have to be produced and distributed under full GMP/GDP regime, including GMP-compliant analyses with validated methods.

**Signatures:** controlled & released:

**Staff:**


**completed:**  
14.04.2016

The enclosed results refer exclusively to the object of examination described above.  
The measuring accuracy is available on request. Only handwritten signatures are valid.  
Without written permission, it is not allowed to publish single parts of this report.

page 1 of 1

**Report steroid analysis**

HPLC:	Nr.5	Druckdatum:	11.04.2016 08:26:09
Analysendatum:	09.04.2016 17:54:38	Injektvolumen:	20
Vial:	32		
Probenidentifikation:	C42475 A		
Operator:	admin		
Chrom.-Beding.	ACN / Wasser, gradient		
Temperatur:	40 °C		
Säule	Nr. 121, 250 x 4,6, C18		
Detektion:	UV 240nm		



**Channel A**

**UV Results**


Name	Time	Area	Amount	Area Percent
Fluoxymesteron	16,310	104250646	0,000 BDL	73,264
Trenbolone enanthate	26,597	38043771	0,000	26,736
<b>Totals</b>		<b>142294417</b>	<b>0,000</b>	<b>100,000</b>

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\05,04,16\045\_C42475 A-Rep1.dat  
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

## Report steroid analysis

HPLC: Nr.5  
 Analysendatum: 09.04.2016 18:48:07  
 Vial: 32  
 Probenidentifikation: C42475 A  
 Operator: admin  
 Chrom.-Beding.: ACN / Wasser, gradient  
 Temperatur: 40 °C  
 Säule: Nr. 121, 250 x 4,6, C18  
 Detektion: UV 240nm

Druckdatum: 11.04.2016 08:26:13  
 Injektvolumen: 20



## Channel A


UV Results				
Name	Time	Area	Amount	Area Percent
Fluoxymesteron			0,000 BDL	
	16,327	104670759	0,000	73,336
Trenbolone enanthate	26,623	38057103	0,000	26,664
<b>Totals</b>		<b>142727862</b>	<b>0,000</b>	<b>100,000</b>

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse  
 Steroide\05,04,16\045\_C42475 A-Rep2.dat  
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC: Nr.5  
 Analysendatum: 09.04.2016 19:41:37  
 Vial: 32  
 Probenidentifikation: C42475 A  
 Operator: admin  
 Chrom.-Beding. ACN / Wasser, gradient  
 Temperatur: 40 °C  
 Säule Nr. 121, 250 x 4,6, C18  
 Detektion: UV 240nm

Druckdatum: 11.04.2016 08:26:17  
 Injektvolumen: 20



Channel A


UV Results

Name	Time	Area	Amount	Area Percent
Fluoxymesteron	16,360	104668850	0,000 BDL	73,379
Trenbolone enanthate	26,593	37972109	0,000	26,621
<b>Totals</b>		<b>142640959</b>	<b>0,000</b>	<b>100,000</b>

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\05,04,16\045\_C42475 A-Rep3.dat  
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

**Report steroid analysis**

HPLC: Nr.5  
 Analysendatum: 09.04.2016 20:35:06 Druckdatum: 11.04.2016 08:26:21  
 Vial: 33 Injektvolumen: 20  
 Probenidentifikation: C42475 B  
 Operator: admin  
 Chrom.-Beding. ACN / Wasser, gradient  
 Temperatur: 40 °C  
 Säule Nr. 121, 250 x 4,6, C18  
 Detektion: UV 240nm




**Channel A**

UV Results				
Name	Time	Area	Amount	Area Percent
Fluoxymesteron			0,000 BDL	
	16,360	105183741	0,000	73,309
Trenbolone enanthate	26,623	38296245	0,000	26,691
Totals		143479986	0,000	100,000

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\05,04,16\046\_C42475 B-Rep1.dat  
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

**Report steroid analysis**

HPLC:	Nr.5	Druckdatum:	11.04.2016 08:26:24
Analysendatum:	09.04.2016 21:28:30	Injektvolumen:	20
Vial:	33		
Probenidentifikation:	C42475 B		
Operator:	admin		
Chrom.-Beding.	ACN / Wasser, gradient		
Temperatur:	40 °C		
Säule	Nr. 121, 250 x 4,6, C18		
Detektion:	UV 240nm		



**Channel A**


UV Results				
Name	Time	Area	Amount	Area Percent
Fluoxymesteron			0,000 BDL	
	16,357	105347223	0,000	73,331
Trenbolone enanthate	26,607	38313448	0,000	26,669
Totals		143660671	0,000	100,000

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\05,04,16\046\_C42475 B-Rep2.dat  
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC: Nr.5  
 Analysendatum: 09.04.2016 22:22:01  
 Vial: 33  
 Probenidentifikation: C42475 B  
 Operator: admin  
 Chrom.-Beding. ACN / Wasser, gradient  
 Temperatur: 40 °C  
 Säule Nr. 121, 250 x 4,6, C18  
 Detektion: UV 240nm

Druckdatum: 11.04.2016 08:26:28  
 Injektvolumen: 20



Channel A

UV Results				
Name	Time	Area	Amount	Area Percent
Fluoxymesteron			0,000 BDL	
	16,307	106110003	0,000	73,389
Trenbolone enanthate	26,597	38474924	0,000	26,611
Totals		144584927	0,000	100,000


Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\05,04,16\046\_C42475 B-Rep3.dat  
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met



**Report steroid analysis**

HPLC: Nr.5  
 Analysendatum: 09.04.2016 23:15:23  
 Vial: 1  
 Probenidentifikation: Blank  
 Operator: admin  
 Chrom.-Beding.: ACN / Wasser, gradient  
 Temperatur: 40 °C  
 Säule: Nr. 121, 250 x 4,6, C18  
 Detektion: UV 240nm

Druckdatum: 11.04.2016 08:26:32  
 Injektvolumen: 20



**Channel A**

UV Results

Name	Time	Area	Amount	Area Percent
Fluoxymesteron			0,000 BDL	
Trenbolone enanthate			0,000 BDL	

Totals				
--------	--	--	--	--

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\05,04,16\047\_Blank.dat  
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met